

NANOSPHERE™ SIZE STANDARDS NIST Traceable Mean Diameter

1. DESCRIPTION. These particle size standards provide accurate and traceable size calibration for particle size analysis. They are part of a series of polymer microspheres with calibrated mean diameters traceable to the Standard Meter through the National Institute of Standards and Technology (NIST). Diameters from 20 nanometers (nm) to 160 micrometers (μm) are available as aqueous suspensions in dropper-tipped vials, calibrated by photon correlation spectroscopy (PCS), transmission electron microscopy (TEM) or optical microscopy. The aqueous medium has been prepared to promote dispersion and reduce clumping of the particles. The approximate particle concentration in percent solids is given to facilitate dilution for the calibration and validation of particle analyzers. Diameters from 200 μm to 1000 μm are available as dry spheres, calibrated by optical microscopy. The certified mean diameter is traceable to NIST. Other values are for information only and should not be used as calibration values.

2. PHYSICAL DATA.

Certified Mean Diameter:	510 nm \pm 7 nm, k=2	Catalog Number: 3500 and 3500A, Nominal 500 nm
Standard Deviation:	9.2 nm	
Coefficient of Variation:	1.8%	
Microsphere Composition:	Polystyrene	
Microsphere Density:	1.05 g/cm ³	
Index of Refraction:	1.59 @ 589 nm	
Approximate Concentration:	1% solids	

- Continued on page 2

CERTIFICATE OF CALIBRATION AND TRACEABILITY

This certifies that the calibrated mean diameter was transferred by transmission electron microscopy (TEM) from the National Institute of Standards and Technology (NIST) certified microspheres (Standard Reference Material 1963, 1691 or 1690).

Catalog Number: 3500 and 3500A, Nanosphere™ Size Standards

Certification Date:	March 21, 2019
Certified Batch:	3500-005
Production Batch:	3500-035
Certified Mean Diameter:	510 nm
Expanded Uncertainty:	\pm 7 nm, k=2





Jennifer Liu, Scientist II
Thermo Fisher Scientific Particle Technology

12/03/19

Packaging Lot # 238846

Expiration Date: MAR'24